

## INSPECTION CERTIFICATE

Cert No: **DIF/A-C- 087/20**

Applicant / Certificate Holder	<b>Ydropsiktiki Papathanasiou SA,</b> 1st Km.Markopoulou – Kalyvion ave. Markopoulo Mesogeion - 19003
Date of Application:	23/12/2020
Product description / Intended Use:	Air sterilization device through UV lamps, in an aluminium casing, floor standing.
Type / S/N of prototype :	<b>Floor devices,</b> <b>Series PURE 60-100-140/PRO, PURE ROOM /</b> <b>s/n: 140.00192, PURE 140/PRO</b>
Specifications Reference:	<b>EU Directive 25/2006, EN 14255-1:2005</b>
Inspection - Testing Facilities:	Ydropsiktiki Papathanasiou SA, Design and construction facilities 1st Km.Markopoulou – Kalyvion ave. Markopoulo Mesogeion - 19003
Inspection Report, No./ Date :	<b>DIF/A-R-087/20 / 30.12.2020</b>
Attachment, Certificate Documents:	Manufacturer's technical File Measurement results, ( SL02EBE2020/03 )

EBETAM SA after performing measurements on a sample of PURE 140/PRO, with four lamps, OSRAM 4x25W, type (HNS 25 W G13), certifies that based on the results reported in the inspection report, there is Compliance with the provisions / requirements of the mentioned directive and Norm, regarding Observance of the safe limit of maximum Ultraviolet Dose, (Exposure Limit Value).

The Certificate concerns and is valid only for the tested type of product, for floor use, as evaluated by EBETAM SA. at the specific place of Inspection.

**Validity:** until December 2023

**Place/Date of Issue:** Athens 31.12.2020

**For EBETAM SA.**

I. Dimitriadis  
Dipl. Electrical Engineer  
Director of the Athens Office



AET : 36413

Γραφείο Προέδρου Αθήνα: Δραγατσανίου 6, 105 59 Αθήνα • T 216 0016999 • E n.sofianos@mirtec.gr  
Έδρα: Α' Βιομηχανική Περιοχή Βόλου, Τ.Θ. 13, 385 00 Βόλος • T 24213 13500 • E volos.office@mirtec.gr  
Παράρτημα Θήβας: 76ο χλμ Ε.Ο. Αθηνών - Λαμίας, Τ.Θ. 150, 320 09 Σχηματάρι • T 22620 71811 • E thiva.office@mirtec.gr  
Παράρτημα Καλλιθέας: Ελ. Βενιζέλου 4, 176 76 Καλλιθέα • T 210 9234932 • E athens.branch@mirtec.gr  
Γραφείο Αθηνών: Μελίνας Μερκούρη 76, 173 42 Αθήνα • T 210 9961408 • E athens.office@mirtec.gr  
Γραφείο Θεσσαλονίκης: Βιομηχανική Περιοχή, 570 22 Θεσσαλονίκη • T 23107 97887 • E thess.office@mirtec.gr